

PRATIKA 100-05

Bilancia elettronica fino a 100 kg

Codice	Modello	Dimensioni	Peso
09013019	PRATIKA 100-05	380 x 300 x 80 mm	3,10 kg



Caratteristiche tecniche

Capacità massima	100 kg
Risoluzione	5 g
Precisione	±0,5 % del valore
Dimensione piattaforma	230 x 230 mm
Unità di misura	kg, lbs/oz
Temperatura di utilizzo	0° - 45° C
Autospegnimento	1.5 min
Alimentazione	batteria 9V alcalina



Taratura

Codice	Modello	Descrizione
09053013		Rapporto taratura rif. Accredia per bilancia
09013019001	PRATIKA 100-05/RT	Bilancia completa di rapporto taratura



PRATIKA-PLUS 100-05

Bilancia elettronica programmabile fino a 100 kg

Codice	Modello	Dimensioni	Peso
09013020	PRATIKA-PLUS 100-05	380 x 300 x 80 mm	3,70 kg



Caratteristiche tecniche

Capacità massima	100 kg
Risoluzione	5 g
Precisione	±0,5 % del valore
Dimensione piattaforma	230 x 230 mm
Unità di misura	kg, lbs/oz
Temperatura di utilizzo	0° - 45° C
Autospegnimento	1.5 min
Alimentazione	4x1,5 AA



Taratura

Codice	Modello	Descrizione
09053013		Rapporto taratura rif. Accredia per bilancia
09013020001	PRATIKA-PLUS 100-05/RT	Bilancia completa di rapporto taratura



PRATIKA 100-05

Electronic scale up to 100 kg

Code	Model	Dimensions	Weight
09013019	PRATIKA 100-05	380 x 300 x 80 mm	3,10 kg



Technical features

Max capacity	100 kg
Resolution	5 g
Accuracy	0,5 % value
Scale pan dimensions	230 x 230 mm
Measuring units	kg, lbs/oz
Operating temperature	0° - 45° C
Auto power off	1.5 min
Power supply	9V alkaline batteries



Calibration

Code	Model	Description
09053013		Calibration report ref. Accredia for scale
09013019001	PRATIKA 100-05/RT	Scale with calibration report



PRATIKA-PLUS 100-05

Programmable electronic scale up to 100 kg

Code	Model	Dimensions	Weight
09013020	PRATIKA-PLUS 100-05	380 x 300 x 80 mm	3,70 kg



Technical features

Max capacity	100 kg
Resolution	5 g
Accuracy	0,5 % value
Scale pan dimensions	230 x 230 mm
Measuring units	kg, lbs/oz
Operating temperature	0° - 45° C
Auto power off	1.5 min
Power supply	4x1,5 AA



Calibration

Code	Model	Description
09053013		Calibration report ref. Accredia for scale
09013020001	PRATIKA-PLUS 100-05/RT	Scale with calibration report





Wigam S.p.A.
Loc.Spedale 10/b
52018 Castel San Niccolò
(AR) Italia
Tel. 0575 / 5011
Fax. 0575 / 501200



Dichiarazione di conformità

Noi, firmatari della presente, dichiariamo sotto la nostra esclusiva responsabilità, che la bilancia, modello :

PRATIKA 100

è progettata secondo quanto prescritto dalla direttiva :

- 2014/30/UE Direttiva sulla compatibilità elettromagnetica

Test standard:

- EN 55022 :2006+A1 :2007
- EN 55024:1998+A1:2001+A2:2003

Castel San Niccolò
11/05/2016

Gastone Vangelisti
(Presidente)



Wigam S.p.A.
Loc.Spedale 10/b
52018 Castel San Niccolò
(AR) Italia
Tel. 0575 / 5011
Fax. 0575 / 501200



Declaration of conformity

We, signers of this declaration, declare under our own exclusive responsibility, that the scale, model :

PRATIKA 100

Are in compliance with the following Standards:

- 2014/30/UE Directive on electromagnetic compatibility

Used test standards

- EN 55022 :1998+A2 :2003
- EN 55024:1998+A2:2003

Castel San Niccolò
11/05/2016

Gastone Vangelisti
(President)