

**PRATIKA 100-05**

Bilancia elettronica fino a 100 kg

Codice	Modello	Dimensioni	Peso
09013019	PRATIKA 100-05	380 x 300 x 80 mm	3,10 kg


**Caratteristiche tecniche**

Capacità massima	100 kg
Risoluzione	5 g
Precisione	±0,5 % del valore
Dimensione piattaforma	230 x 230 mm
Unità di misura	kg, lbs/oz
Temperatura di utilizzo	0° - 45° C
Autospegnimento	1.5 min
Alimentazione	batteria 9V alcalina


**Taratura**

Codice	Modello	Descrizione
09053013		Rapporto taratura rif. Accredia per bilancia
09013019001	PRATIKA 100-05/RT	Bilancia completa di rapporto taratura


**PRATIKA-PLUS 100-05**

Bilancia elettronica programmabile fino a 100 kg

Codice	Modello	Dimensioni	Peso
09013020	PRATIKA-PLUS 100-05	380 x 300 x 80 mm	3,70 kg


**Caratteristiche tecniche**

Capacità massima	100 kg
Risoluzione	5 g
Precisione	±0,5 % del valore
Dimensione piattaforma	230 x 230 mm
Unità di misura	kg, lbs/oz
Temperatura di utilizzo	0° - 45° C
Autospegnimento	1.5 min
Alimentazione	4x1,5 AA


**Taratura**

Codice	Modello	Descrizione
09053013		Rapporto taratura rif. Accredia per bilancia
09013020001	PRATIKA-PLUS 100-05/RT	Bilancia completa di rapporto taratura



**PRATIKA 100-05**

Electronic scale up to 100 kg

Code	Model	Dimensions	Weight
09013019	PRATIKA 100-05	380 x 300 x 80 mm	3,10 kg


**Technical features**

Max capacity	100 kg
Resolution	5 g
Accuracy	0,5 % value
Scale pan dimensions	230 x 230 mm
Measuring units	kg, lbs/oz
Operating temperature	0° - 45° C
Auto power off	1.5 min
Power supply	9V alkaline batteries


**Calibration**

Code	Model	Description
09053013		Calibration report ref. Accredia for scale
09013019001	PRATIKA 100-05/RT	Scale with calibration report

**PRATIKA-PLUS 100-05**

Programmable electronic scale up to 100 kg

Code	Model	Dimensions	Weight
09013020	PRATIKA-PLUS 100-05	380 x 300 x 80 mm	3,70 kg


**Technical features**

Max capacity	100 kg
Resolution	5 g
Accuracy	0,5 % value
Scale pan dimensions	230 x 230 mm
Measuring units	kg, lbs/oz
Operating temperature	0° - 45° C
Auto power off	1.5 min
Power supply	4x1,5 AA


**Calibration**

Code	Model	Description
09053013		Calibration report ref. Accredia for scale
09013020001	PRATIKA-PLUS 100-05/RT	Scale with calibration report



Wigam S.p.A.  
Loc.Spedale 10/b  
52018 Castel San Niccolò  
(AR) Italia  
Tel. 0575 / 5011  
Fax. 0575 / 501200



## Dichiarazione di conformità

Noi, firmatari della presente, dichiariamo sotto la nostra esclusiva responsabilità, che la bilancia, modello :

### ***PRATIKA 100***

è progettata secondo quanto prescritto dalla direttiva :

- 2014/30/UE Direttiva sulla compatibilità elettromagnetica

Test standard:

- EN 55022 :2006+A1 :2007
- EN 55024:1998+A1:2001+A2:2003

Castel San Niccolò  
11/05/2016

Gastone Vangelisti  
(Presidente)

A handwritten signature in black ink, appearing to read 'Gastone Vangelisti'.



Wigam S.p.A.  
Loc.Spedale 10/b  
52018 Castel San Niccolò  
(AR) Italia  
Tel. 0575 / 5011  
Fax. 0575 / 501200

---



## Declaration of conformity

---

We, signers of this declaration, declare under our own exclusive responsibility, that the scale, model :

### **PRATIKA 100**

Are in compliance with the following Standards:

- 2014/30/UE Directive on electromagnetic compatibility

Used test standards

- EN 55022 :1998+A2 :2003
- EN 55024:1998+A2:2003

Castel San Niccolò  
11/05/2016

Gastone Vangelisti  
(President)

A handwritten signature in black ink, appearing to read "Gastone Vangelisti".